

DIONEX CAPILLARY ELECTROPHORESIS INSTRUMENT

This instrument shall be calibrated before each use in casework as outlined by the procedures below.

An anion and/or cation standard will be prepared in accordance with the “Standard Operating Procedures.” Each anion or cation should be present to ensure that the instrument is operating correctly. Concentrations of ions should be similar to previous calibrations. The date, operator, and case number shall be recorded on the calibration log.