

## **MUFFLE FURNACE CALIBRATION**

-Calibration of the muffle furnace will be done before each use.

-Procedure

1. Record Date
2. Read the furnace setting and record on the muffle furnace calibration log.
3. Read the internal thermometer and record the temperature in the calibration log.
4. Initials of the chemist performing the calibration should be recorded beside the results in the calibration log.
5. If calibration procedure shows the instrument to be inconsistent it shall be serviced by a qualified service engineer.

